

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/06/2024

Lab Sample No.: AR1660CCC500 Data File : PO107726.D Time Analyzed: 09:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.423	5.623	511.290	500.000	2.3
Aroclor-1016-2	5.547	5.445	5.645	497.460	500.000	-0.5
Aroclor-1016-3	5.610	5.507	5.707	489.050	500.000	-2.2
Aroclor-1016-4	5.707	5.604	5.804	512.950	500.000	2.6
Aroclor-1016-5	6.002	5.899	6.099	511.100	500.000	2.2
Aroclor-1260-1	7.132	7.026	7.226	513.000	500.000	2.6
Aroclor-1260-2	7.389	7.282	7.482	542.700	500.000	8.5
Aroclor-1260-3	7.752	7.643	7.843	537.760	500.000	7.6
Aroclor-1260-4	7.977	7.868	8.068	552.670	500.000	10.5
Aroclor-1260-5	8.291	8.181	8.381	599.200	500.000	19.8
Decachlorobiphenyl	10.079	9.958	10.158	48.100	50.000	-3.8
Tetrachloro-m-xylene	4.373	4.274	4.474	51.680	50.000	3.4

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/06/2024

Lab Sample No.: AR1660CCC500 Data File : PO107726.D Time Analyzed: 09:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.724	4.626	4.826	535.170	500.000	7.0
Aroclor-1016-2	4.743	4.645	4.845	548.800	500.000	9.8
Aroclor-1016-3	4.918	4.821	5.021	527.740	500.000	5.5
Aroclor-1016-4	4.960	4.862	5.062	488.290	500.000	-2.3
Aroclor-1016-5	5.174	5.076	5.276	523.130	500.000	4.6
Aroclor-1260-1	6.204	6.107	6.307	506.050	500.000	1.2
Aroclor-1260-2	6.392	6.294	6.494	533.470	500.000	6.7
Aroclor-1260-3	6.545	6.448	6.648	522.150	500.000	4.4
Aroclor-1260-4	7.017	6.918	7.118	519.510	500.000	3.9
Aroclor-1260-5	7.258	7.159	7.359	552.050	500.000	10.4
Decachlorobiphenyl	8.640	8.539	8.739	54.490	50.000	9.0
Tetrachloro-m-xylene	3.643	3.545	3.745	54.140	50.000	8.3