

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/06/2024

Lab Sample No.: AR1660CCC500 Data File : PO107741.D Time Analyzed: 16:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.524	5.423	5.623	463.130	500.000	-7.4
Aroclor-1016-2	5.546	5.445	5.645	444.160	500.000	-11.2
Aroclor-1016-3	5.608	5.507	5.707	441.970	500.000	-11.6
Aroclor-1016-4	5.706	5.604	5.804	462.710	500.000	-7.5
Aroclor-1016-5	6.001	5.899	6.099	463.090	500.000	-7.4
Aroclor-1260-1	7.130	7.026	7.226	456.620	500.000	-8.7
Aroclor-1260-2	7.387	7.282	7.482	464.280	500.000	-7.1
Aroclor-1260-3	7.750	7.643	7.843	448.010	500.000	-10.4
Aroclor-1260-4	7.975	7.868	8.068	465.650	500.000	-6.9
Aroclor-1260-5	8.289	8.181	8.381	482.090	500.000	-3.6
Decachlorobiphenyl	10.074	9.958	10.158	43.960	50.000	-12.1
Tetrachloro-m-xylene	4.372	4.274	4.474	52.670	50.000	5.3

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/06/2024

Lab Sample No.: AR1660CCC500 Data File : PO107741.D Time Analyzed: 16:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.626	4.826	534.360	500.000	6.9
Aroclor-1016-2	4.742	4.645	4.845	554.800	500.000	11.0
Aroclor-1016-3	4.918	4.821	5.021	533.240	500.000	6.6
Aroclor-1016-4	4.959	4.862	5.062	489.970	500.000	-2.0
Aroclor-1016-5	5.173	5.076	5.276	523.370	500.000	4.7
Aroclor-1260-1	6.204	6.107	6.307	496.980	500.000	-0.6
Aroclor-1260-2	6.391	6.294	6.494	518.510	500.000	3.7
Aroclor-1260-3	6.544	6.448	6.648	492.660	500.000	-1.5
Aroclor-1260-4	7.016	6.918	7.118	462.170	500.000	-7.6
Aroclor-1260-5	7.256	7.159	7.359	474.270	500.000	-5.1
Decachlorobiphenyl	8.637	8.539	8.739	41.590	50.000	-16.8
Tetrachloro-m-xylene	3.643	3.545	3.745	53.650	50.000	7.3