

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 11/06/2024

Lab Sample No.: AR1660CCC500 Data File : PO107756.D Time Analyzed: 21:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.523	5.423	5.623	521.970	500.000	4.4
Aroclor-1016-2	5.546	5.445	5.645	513.220	500.000	2.6
Aroclor-1016-3	5.608	5.507	5.707	503.460	500.000	0.7
Aroclor-1016-4	5.705	5.604	5.804	530.050	500.000	6.0
Aroclor-1016-5	6.001	5.899	6.099	533.120	500.000	6.6
Aroclor-1260-1	7.132	7.026	7.226	487.030	500.000	-2.6
Aroclor-1260-2	7.388	7.282	7.482	514.800	500.000	3.0
Aroclor-1260-3	7.751	7.643	7.843	532.710	500.000	6.5
Aroclor-1260-4	7.976	7.868	8.068	568.210	500.000	13.6
Aroclor-1260-5	8.290	8.181	8.381	561.040	500.000	12.2
Decachlorobiphenyl	10.077	9.958	10.158	47.560	50.000	-4.9
Tetrachloro-m-xylene	4.372	4.274	4.474	52.470	50.000	4.9

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 11/06/2024

Lab Sample No.: AR1660CCC500 Data File : PO107756.D Time Analyzed: 21:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.724	4.626	4.826	531.640	500.000	6.3
Aroclor-1016-2	4.743	4.645	4.845	551.000	500.000	10.2
Aroclor-1016-3	4.919	4.821	5.021	530.220	500.000	6.0
Aroclor-1016-4	4.960	4.862	5.062	490.360	500.000	-1.9
Aroclor-1016-5	5.174	5.076	5.276	523.760	500.000	4.8
Aroclor-1260-1	6.205	6.107	6.307	511.170	500.000	2.2
Aroclor-1260-2	6.392	6.294	6.494	542.160	500.000	8.4
Aroclor-1260-3	6.546	6.448	6.648	529.960	500.000	6.0
Aroclor-1260-4	7.017	6.918	7.118	530.010	500.000	6.0
Aroclor-1260-5	7.258	7.159	7.359	565.090	500.000	13.0
Decachlorobiphenyl	8.639	8.539	8.739	54.600	50.000	9.2
Tetrachloro-m-xylene	3.643	3.545	3.745	53.910	50.000	7.8