

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL04 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107763.D Time Analyzed: 00:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.523	5.423	5.623	547.580	500.000	9.5
Aroclor-1016-2	5.546	5.445	5.645	529.750	500.000	6.0
Aroclor-1016-3	5.608	5.507	5.707	527.770	500.000	5.6
Aroclor-1016-4	5.706	5.604	5.804	552.330	500.000	10.5
Aroclor-1016-5	6.001	5.899	6.099	549.630	500.000	9.9
Aroclor-1260-1	7.131	7.026	7.226	525.580	500.000	5.1
Aroclor-1260-2	7.388	7.282	7.482	540.410	500.000	8.1
Aroclor-1260-3	7.750	7.643	7.843	537.840	500.000	7.6
Aroclor-1260-4	7.976	7.868	8.068	568.050	500.000	13.6
Aroclor-1260-5	8.290	8.181	8.381	579.820	500.000	16.0
Decachlorobiphenyl	10.077	9.958	10.158	48.730	50.000	-2.5
Tetrachloro-m-xylene	4.372	4.274	4.474	53.790	50.000	7.6

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL04 Date Analyzed: 11/07/2024

Lab Sample No.: AR1660CCC500 Data File : PO107763.D Time Analyzed: 00:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.626	4.826	549.090	500.000	9.8
Aroclor-1016-2	4.742	4.645	4.845	562.330	500.000	12.5
Aroclor-1016-3	4.918	4.821	5.021	540.540	500.000	8.1
Aroclor-1016-4	4.960	4.862	5.062	495.890	500.000	-0.8
Aroclor-1016-5	5.173	5.076	5.276	536.380	500.000	7.3
Aroclor-1260-1	6.204	6.107	6.307	525.020	500.000	5.0
Aroclor-1260-2	6.392	6.294	6.494	553.370	500.000	10.7
Aroclor-1260-3	6.545	6.448	6.648	536.080	500.000	7.2
Aroclor-1260-4	7.016	6.918	7.118	536.010	500.000	7.2
Aroclor-1260-5	7.257	7.159	7.359	565.900	500.000	13.2
Decachlorobiphenyl	8.639	8.539	8.739	55.530	50.000	11.1
Tetrachloro-m-xylene	3.642	3.545	3.745	55.140	50.000	10.3