

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028451.D Time Analyzed: 13:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.254	9.160	9.360	735.550	712.500	3.2
2,4-D	8.374	8.279	8.479	712.300	705.000	1.0
2,4-DCAA	7.244	7.150	7.350	740.980	750.000	-1.2

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028451.D Time Analyzed: 13:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.902	9.811	10.011	664.220	712.500	-6.8
2,4-D	8.995	8.905	9.105	639.620	705.000	-9.3
2,4-DCAA	7.747	7.655	7.855	681.430	750.000	-9.1