

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/11/2024

Lab Sample No.: PSTDCCC050 Data File : PL092943.D Time Analyzed: 11:58

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|----------------------|-------|--------------------|-------------------|-------|
| Decachlorobiphenyl   | 9.061 | 8.954                | 9.154 | 46.430             | 50.000            | -7.1  |
| Endrin               | 6.579 | 6.474                | 6.674 | 43.850             | 50.000            | -12.3 |
| gamma-BHC (Lindane)  | 4.331 | 4.228                | 4.428 | 49.900             | 50.000            | -0.2  |
| Heptachlor           | 4.920 | 4.817                | 5.017 | 47.730             | 50.000            | -4.5  |
| Heptachlor epoxide   | 5.688 | 5.584                | 5.784 | 48.400             | 50.000            | -3.2  |
| Methoxychlor         | 7.504 | 7.399                | 7.599 | 44.210             | 50.000            | -11.6 |
| Tetrachloro-m-xylene | 3.542 | 3.440                | 3.640 | 49.960             | 50.000            | -0.1  |

### CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/11/2024

Lab Sample No.: PSTDCCC050 Data File : PL092943.D Time Analyzed: 11:58

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Decachlorobiphenyl   | 7.919 | 7.816                | 8.016 | 46.770             | 50.000            | -6.5 |
| Endrin               | 5.643 | 5.541                | 5.741 | 50.950             | 50.000            | 1.9  |
| gamma-BHC (Lindane)  | 3.611 | 3.511                | 3.711 | 52.560             | 50.000            | 5.1  |
| Heptachlor           | 3.951 | 3.849                | 4.049 | 50.950             | 50.000            | 1.9  |
| Heptachlor epoxide   | 4.733 | 4.632                | 4.832 | 52.470             | 50.000            | 4.9  |
| Methoxychlor         | 6.617 | 6.515                | 6.715 | 46.930             | 50.000            | -6.1 |
| Tetrachloro-m-xylene | 2.778 | 2.678                | 2.878 | 51.750             | 50.000            | 3.5  |