

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/11/2024

Lab Sample No.: PSTDCCC050 Data File : PL092960.D Time Analyzed: 16:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.064	8.954	9.154	45.750	50.000	-8.5
Endrin	6.580	6.474	6.674	43.680	50.000	-12.6
gamma-BHC (Lindane)	4.332	4.228	4.428	49.740	50.000	-0.5
Heptachlor	4.921	4.817	5.017	46.230	50.000	-7.5
Heptachlor epoxide	5.689	5.584	5.784	48.040	50.000	-3.9
Methoxychlor	7.506	7.399	7.599	41.600	50.000	-16.8
Tetrachloro-m-xylene	3.543	3.440	3.640	50.620	50.000	1.2

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/11/2024

Lab Sample No.: PSTDCCC050 Data File : PL092960.D Time Analyzed: 16:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.920	7.816	8.016	47.520	50.000	-5.0
Endrin	5.643	5.541	5.741	51.290	50.000	2.6
gamma-BHC (Lindane)	3.611	3.511	3.711	52.740	50.000	5.5
Heptachlor	3.950	3.849	4.049	49.830	50.000	-0.3
Heptachlor epoxide	4.733	4.632	4.832	52.930	50.000	5.9
Methoxychlor	6.617	6.515	6.715	45.570	50.000	-8.9
Tetrachloro-m-xylene	2.779	2.678	2.878	51.800	50.000	3.6