

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL04 Date Analyzed: 11/12/2024

Lab Sample No.: PSTDCCC050 Data File : PL092993.D Time Analyzed: 01:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.062	8.954	9.154	45.740	50.000	-8.5
Endrin	6.579	6.474	6.674	42.020	50.000	-16.0
gamma-BHC (Lindane)	4.331	4.228	4.428	49.360	50.000	-1.3
Heptachlor	4.919	4.817	5.017	45.840	50.000	-8.3
Heptachlor epoxide	5.686	5.584	5.784	47.100	50.000	-5.8
Methoxychlor	7.504	7.399	7.599	40.210	50.000	-19.6
Tetrachloro-m-xylene	3.542	3.440	3.640	50.220	50.000	0.4

CALIBRATION VERIFICATION SUMMARY

Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL04 Date Analyzed: 11/12/2024

Lab Sample No.: PSTDCCC050 Data File : PL092993.D Time Analyzed: 01:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.920	7.816	8.016	46.990	50.000	-6.0
Endrin	5.643	5.541	5.741	48.620	50.000	-2.8
gamma-BHC (Lindane)	3.611	3.511	3.711	52.480	50.000	5.0
Heptachlor	3.950	3.849	4.049	49.220	50.000	-1.6
Heptachlor epoxide	4.733	4.632	4.832	51.790	50.000	3.6
Methoxychlor	6.617	6.515	6.715	42.960	50.000	-14.1
Tetrachloro-m-xylene	2.778	2.678	2.878	51.720	50.000	3.4