

Surrogate Summary

SW-846

SDG No.: P4718

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4718-01	WB-307-SB01	2-Fluorophenol	150	135	90		30 (18)	130 (112)
		Phenol-d6	150	134	90		30 (15)	130 (107)
		Nitrobenzene-d5	100	90.3	90		30 (18)	130 (107)
		2-Fluorobiphenyl	100	82.9	83		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	151	101		30 (10)	130 (116)
		Terphenyl-d14	100	80.7	81		30 (10)	130 (105)
P4718-02	WB-307-SB02	2-Fluorophenol	150	119	79		30 (18)	130 (112)
		Phenol-d6	150	119	79		30 (15)	130 (107)
		Nitrobenzene-d5	100	78.5	78		30 (18)	130 (107)
		2-Fluorobiphenyl	100	78.9	79		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	134	89		30 (10)	130 (116)
		Terphenyl-d14	100	81.5	81		30 (10)	130 (105)
P4718-02MS	WB-307-SB02MS	2-Fluorophenol	150	108	72		30 (18)	130 (112)
		Phenol-d6	150	109	73		30 (15)	130 (107)
		Nitrobenzene-d5	100	71.6	72		30 (18)	130 (107)
		2-Fluorobiphenyl	100	71.9	72		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	121	81		30 (10)	130 (116)
		Terphenyl-d14	100	80.3	80		30 (10)	130 (105)
P4718-02MSD	WB-307-SB02MSD	2-Fluorophenol	150	106	70		30 (18)	130 (112)
		Phenol-d6	150	106	71		30 (15)	130 (107)
		Nitrobenzene-d5	100	69.7	70		30 (18)	130 (107)
		2-Fluorobiphenyl	100	71.3	71		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	121	81		30 (10)	130 (116)
		Terphenyl-d14	100	76.8	77		30 (10)	130 (105)
PB164702BL	PB164702BL	2-Fluorophenol	150	137	91		30 (18)	130 (112)
		Phenol-d6	150	134	89		30 (15)	130 (107)
		Nitrobenzene-d5	100	91.0	91		30 (18)	130 (107)
		2-Fluorobiphenyl	100	91.1	91		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	152	101		30 (10)	130 (116)
		Terphenyl-d14	100	101	101		30 (10)	130 (105)
PB164702BS	PB164702BS	2-Fluorophenol	150	125	83		30 (18)	130 (112)
		Phenol-d6	150	122	82		30 (15)	130 (107)
		Nitrobenzene-d5	100	86.5	86		30 (18)	130 (107)
		2-Fluorobiphenyl	100	86.4	86		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	149	99		30 (10)	130 (116)
		Terphenyl-d14	100	97.3	97		30 (10)	130 (105)

Surrogate Summary

SW-846

SDG No.: P4718

Client: Portal Partners Tri-Venture

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P4718-04	WB-307-SW	2-Fluorophenol	150	35.2	23		15 (10)	110 (139)
		Phenol-d6	150	30.7	20		15 (10)	110 (134)
		Nitrobenzene-d5	100	35.6	36		30 (49)	130 (133)
		2-Fluorobiphenyl	100	36.4	36		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	48.6	32		15 (44)	110 (137)
		Terphenyl-d14	100	39.4	39		30 (48)	130 (125)
PB164846BL	PB164846BL	2-Fluorophenol	150	135	90		15 (10)	110 (139)
		Phenol-d6	150	130	87		15 (10)	110 (134)
		Nitrobenzene-d5	100	92.4	92		30 (49)	130 (133)
		2-Fluorobiphenyl	100	93.6	94		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	135	90		15 (44)	110 (137)
		Terphenyl-d14	100	85.1	85		30 (48)	130 (125)
PB164846BS	PB164846BS	2-Fluorophenol	150	119	79		15 (10)	110 (139)
		Phenol-d6	150	116	77		15 (10)	110 (134)
		Nitrobenzene-d5	100	81.8	82		30 (49)	130 (133)
		2-Fluorobiphenyl	100	83.2	83		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	129	86		15 (44)	110 (137)
		Terphenyl-d14	100	102	102		30 (48)	130 (125)
PB164846BSD	PB164846BSD	2-Fluorophenol	150	126	84		15 (10)	110 (139)
		Phenol-d6	150	121	81		15 (10)	110 (134)
		Nitrobenzene-d5	100	86.0	86		30 (49)	130 (133)
		2-Fluorobiphenyl	100	87.5	88		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	132	88		15 (44)	110 (137)
		Terphenyl-d14	100	99.7	100		30 (48)	130 (125)