

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4718

Client: Portal Partners Tri-Venture

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P4718-02MSD	Client Sample ID:	WB-307-SB02MSD					DataFile:	BF140273.D		
Benzaldehyde	2100	0	440	ug/Kg	21		0		20 (10)	160 (86)	30 (20)
Phenol	2100	0	2200	ug/Kg	105		5		20 (67)	160 (126)	30 (20)
bis(2-Chloroethyl)ether	2100	0	2200	ug/Kg	105		0		70 (54)	130 (125)	30 (20)
2-Chlorophenol	2100	0	2300	ug/Kg	110		0		70 (79)	130 (107)	30 (20)
2-Methylphenol	2100	0	2200	ug/Kg	105		5		70 (66)	130 (122)	30 (20)
2,2-oxybis(1-Chloropropane)	2100	0	2100	ug/Kg	100		5		70 (65)	130 (110)	30 (20)
Acetophenone	2100	0	2100	ug/Kg	100		0		70 (75)	130 (111)	30 (20)
3+4-Methylphenols	2100	0	2200	ug/Kg	105		0		20 (66)	160 (104)	30 (20)
N-Nitroso-di-n-propylamine	2100	0	2100	ug/Kg	100		5		70 (59)	130 (119)	30 (20)
Hexachloroethane	2100	0	2100	ug/Kg	100		0		20 (65)	160 (117)	30 (20)
Nitrobenzene	2100	0	2000	ug/Kg	95		0		70 (70)	130 (119)	30 (20)
Isophorone	2100	0	2100	ug/Kg	100		5		70 (76)	130 (122)	30 (20)
2-Nitrophenol	2100	0	2300	ug/Kg	110		4		70 (54)	130 (145)	30 (20)
2,4-Dimethylphenol	2100	0	2600	ug/Kg	124		4		70 (44)	130 (135)	30 (20)
bis(2-Chloroethoxy)methane	2100	0	2100	ug/Kg	100		0		70 (68)	130 (112)	30 (20)
2,4-Dichlorophenol	2100	0	2200	ug/Kg	105		0		70 (72)	130 (118)	30 (20)
Naphthalene	2100	0	2100	ug/Kg	100		0		70 (72)	130 (110)	30 (20)
4-Chloroaniline	2100	0	1500	ug/Kg	71		0		70 (10)	130 (91)	30 (20)
Hexachlorobutadiene	2100	0	2000	ug/Kg	95		5		70 (66)	130 (114)	30 (20)
Caprolactam	2100	0	2400	ug/Kg	114		0		20 (51)	160 (134)	30 (20)
4-Chloro-3-methylphenol	2100	0	2200	ug/Kg	105		0		70 (57)	130 (132)	30 (20)
2-Methylnaphthalene	2100	0	2200	ug/Kg	105		0		70 (59)	130 (123)	30 (20)
Hexachlorocyclopentadiene	4300	0	8100	ug/Kg	188	*	5		20 (10)	160 (175)	30 (20)
2,4,6-Trichlorophenol	2100	0	2200	ug/Kg	105		5		70 (72)	130 (117)	30 (20)
2,4,5-Trichlorophenol	2100	0	2200	ug/Kg	105		5		70 (72)	130 (117)	30 (20)
1,1-Biphenyl	2100	0	2100	ug/Kg	100		5		70 (75)	130 (113)	30 (20)
2-Chloronaphthalene	2100	0	2100	ug/Kg	100		0		70 (67)	130 (118)	30 (20)
2-Nitroaniline	2100	0	2200	ug/Kg	105		5		70 (69)	130 (127)	30 (20)
Dimethylphthalate	2100	0	2200	ug/Kg	105		5		70 (70)	130 (113)	30 (20)
Acenaphthylene	2100	0	2300	ug/Kg	110		4		70 (79)	130 (118)	30 (20)
2,6-Dinitrotoluene	2100	0	2200	ug/Kg	105		0		70 (70)	130 (125)	30 (20)
3-Nitroaniline	2100	0	1700	ug/Kg	81		6		70 (30)	130 (99)	30 (20)
Acenaphthene	2100	0	2500	ug/Kg	119		0		70 (70)	130 (121)	30 (20)
2,4-Dinitrophenol	4300	0	5000	ug/Kg	116		2		20 (10)	160 (155)	30 (20)
4-Nitrophenol	4300	0	5300	ug/Kg	123		0		20 (45)	160 (133)	30 (20)
Dibenzofuran	2100	0	2200	ug/Kg	105		5		70 (72)	130 (110)	30 (20)
2,4-Dinitrotoluene	2100	0	2300	ug/Kg	110		4		70 (55)	130 (128)	30 (20)
Diethylphthalate	2100	0	2200	ug/Kg	105		5		70 (70)	130 (112)	30 (20)
4-Chlorophenyl-phenylether	2100	0	2100	ug/Kg	100		5		70 (71)	130 (108)	30 (20)
Fluorene	2100	0	2200	ug/Kg	105		0		70 (68)	130 (116)	30 (20)
4-Nitroaniline	2100	0	2300	ug/Kg	110		4		70 (55)	130 (120)	30 (20)
4,6-Dinitro-2-methylphenol	2100	0	2500	ug/Kg	119		4		70 (10)	130 (160)	30 (20)
N-Nitrosodiphenylamine	2100	0	2200	ug/Kg	105		0		70 (73)	130 (118)	30 (20)
4-Bromophenyl-phenylether	2100	0	2200	ug/Kg	105		0		70 (65)	130 (121)	30 (20)
Hexachlorobenzene	2100	0	2100	ug/Kg	100		5		70 (67)	130 (118)	30 (20)
Atrazine	2100	0	2900	ug/Kg	138	*	4		70 (79)	130 (127)	30 (20)

() = LABORATORY INHOUSE LIMIT

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4718

Client: Portal Partners Tri-Venture

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Pentachlorophenol	4300	0	4700	ug/Kg	109		3		20 (47)	160 (128)	30 (20)
Phenanthrene	2100	0	2200	ug/Kg	105		0		70 (52)	130 (128)	30 (20)
Anthracene	2100	0	2300	ug/Kg	110		0		70 (62)	130 (124)	30 (20)
Carbazole	2100	0	2200	ug/Kg	105		0		70 (59)	130 (119)	30 (20)
Di-n-butylphthalate	2100	0	2300	ug/Kg	110		0		70 (69)	130 (118)	30 (20)
Fluoranthene	2100	0	2100	ug/Kg	100		0		70 (44)	130 (125)	30 (20)
Pyrene	2100	0	2300	ug/Kg	110		4		70 (26)	130 (142)	30 (20)
Butylbenzylphthalate	2100	0	2400	ug/Kg	114		4		70 (64)	130 (126)	30 (20)
3,3-Dichlorobenzidine	2100	0	1900	ug/Kg	90		5		70 (33)	130 (116)	30 (20)
Benzo(a)anthracene	2100	0	2200	ug/Kg	105		5		70 (71)	130 (114)	30 (20)
Chrysene	2100	0	2300	ug/Kg	110		0		70 (57)	130 (121)	30 (20)
bis(2-Ethylhexyl)phthalate	2100	0	2200	ug/Kg	105		5		70 (42)	130 (169)	30 (20)
Di-n-octyl phthalate	2100	0	2200	ug/Kg	105		5		70 (23)	130 (175)	30 (20)
Benzo(b)fluoranthene	2100	0	2100	ug/Kg	100		13		70 (67)	130 (121)	30 (20)
Benzo(k)fluoranthene	2100	0	2300	ug/Kg	110		10		70 (57)	130 (134)	30 (20)
Benzo(a)pyrene	2100	0	2400	ug/Kg	114		4		70 (70)	130 (142)	30 (20)
Indeno(1,2,3-cd)pyrene	2100	0	2100	ug/Kg	100		0		70 (40)	130 (129)	30 (20)
Dibenz(a,h)anthracene	2100	0	2100	ug/Kg	100		0		70 (43)	130 (123)	30 (20)
Benzo(g,h,i)perylene	2100	0	1800	ug/Kg	86		0		70 (24)	130 (125)	30 (20)
1,2,4,5-Tetrachlorobenzene	2100	0	2100	ug/Kg	100		5		70 (69)	130 (124)	30 (20)
1,4-Dioxane	2100	0	2000	ug/Kg	95		0		20 (46)	160 (112)	30 (20)
2,3,4,6-Tetrachlorophenol	2100	0	2400	ug/Kg	114		4		70 (69)	130 (112)	30 (20)