

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/14/2024

Lab File ID: BF140351.D Time Analyzed: 09:42

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	138532	6.881	520272	8.16	291630	9.92
UPPER LIMIT	277064	7.381	1040540	8.657	583260	10.416
LOWER LIMIT	69266	6.381	260136	7.657	145815	9.416
EPA SAMPLE NO.						
01 WB-307-SW	137046	6.87	514015	8.15	280819	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.