

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/14/2024

Lab File ID: BF140366.D Time Analyzed: 17:02

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	130579	6.875	484942	8.16	265027	9.91
UPPER LIMIT	261158	7.375	969884	8.657	530054	10.41
LOWER LIMIT	65289.5	6.375	242471	7.657	132514	9.41
EPA SAMPLE NO.						
01 PB164846BL	127444	6.87	481597	8.15	267971	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.