

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/14/2024

Lab File ID: BF140351.D Time Analyzed: 09:42

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	549011	11.404	316250	14.057	275847	15.557
UPPER LIMIT	1098020	11.904	632500	14.557	551694	16.057
LOWER LIMIT	274506	10.904	158125	13.557	137924	15.057
EPA SAMPLE NO.						
01 WB-307-SW	486241	11.39	241422	14.05	265329	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.