

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: PORT06

Lab Code: CHEM

SAS No.: P4718 SDG NO.: P4718

Lab File ID: BF140350.D

DFTPP Injection Date: 11/14/2024

Instrument ID: BNA\_F

DFTPP Injection Time: 08:47

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	37.6
68	Less than 2.0% of mass 69	0.7 ( 1.7 ) 1
69	Mass 69 relative abundance	38.3
70	Less than 2.0% of mass 69	0.2 ( 0.4 ) 1
127	10.0 - 80.0% of mass 198	48.5
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	27.6
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	13.9
442	Greater than 50% of mass 198	88.9
443	15.0 - 24.0% of mass 442	16 ( 18 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140351.D	11/14/2024	09:42
WB-307-SW	P4718-04	BF140363.D	11/14/2024	15:39