

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110624

Review By	Jagrut	Review On	11/7/2024 11:03:46 AM
Supervise By	mohammad	Supervise On	11/8/2024 3:53:54 AM
SubDirectory	BE110624	HP Acquire Method	BNA_E
		HP Processing Method	be110624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12324,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BE101491.D	6 Nov 2024 11:40		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BE101492.D	6 Nov 2024 13:15	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BE101493.D	6 Nov 2024 13:51		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BE101494.D	6 Nov 2024 14:27	Compound #32,54 removed from 5ppm	RC/JU	Ok,M
5	SSTDICC010	SSTDICC010	BE101495.D	6 Nov 2024 15:03		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BE101496.D	6 Nov 2024 15:39	Compound #32 Kept on LR	RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BE101497.D	6 Nov 2024 16:14	The Calibration is Good For 8270 DOD Except com#77 and will NOT be used for 625.1 Method	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BE101498.D	6 Nov 2024 16:50	Com#77(Benzidine) Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
9	SSTDICC060	SSTDICC060	BE101499.D	6 Nov 2024 17:26		RC/JU	Ok,M
10	SSTDICC080	SSTDICC080	BE101500.D	6 Nov 2024 18:02		RC/JU	Ok,M
11	SSTDICV040	ICVBE110624	BE101501.D	6 Nov 2024 18:41		RC/JU	Ok,M
12	PB164561BL	PB164561BL	BE101502.D	6 Nov 2024 19:16		RC/JU	Ok
13	DFTPP	DFTPP	BE101503.D	6 Nov 2024 19:52		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BE101504.D	6 Nov 2024 20:28		RC/JU	Ok,M
15	PB164561TB	PB164561TB	BE101505.D	6 Nov 2024 21:04		RC/JU	Ok
16	PB164561BS	PB164561BS	BE101506.D	6 Nov 2024 21:40		RC/JU	Ok,M

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110624

Review By	Jagrut	Review On	11/7/2024 11:03:46 AM		
Supervise By	mohammad	Supervise On	11/8/2024 3:53:54 AM		
SubDirectory	BE110624	HP Acquire Method	BNA_E	HP Processing Method	be110624
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12324,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

17	P4660-11	WC-CONCRETE-01-C	BE101507.D	6 Nov 2024 22:15		RC/JU	Ok
18	P4667-04	BP-F-6	BE101508.D	6 Nov 2024 22:51		RC/JU	Ok
19	P4667-08	BP-F-5	BE101509.D	6 Nov 2024 23:27	Internal Standard and Surrogate fail	RC/JU	ReRun
20	P4667-12	TP-10	BE101510.D	7 Nov 2024 00:03	Internal Standard Fail	RC/JU	ReRun
21	P4667-16	BP-F-7	BE101511.D	7 Nov 2024 00:39		RC/JU	Ok
22	P4659-04	MH-2	BE101512.D	7 Nov 2024 1:14		RC/JU	Ok
23	P4679-04	MH-1	BE101513.D	7 Nov 2024 1:50		RC/JU	Ok
24	P4680-04	BP-F26	BE101514.D	7 Nov 2024 2:26	Internal Standrad Fail	RC/JU	ReRun
25	P4680-08	BP-F25	BE101515.D	7 Nov 2024 3:02	Internal Standrad Fail	RC/JU	ReRun
26	P4660-07	WC-WOOD-01-C	BE101516.D	7 Nov 2024 3:38		RC/JU	Ok,M
27	P4711-05	CF-613-COMP-16	BE101517.D	7 Nov 2024 4:13	Internal Standrad Fail	RC/JU	ReRun
28	P4711-10	CF-613-COMP-17	BE101518.D	7 Nov 2024 4:49	Internal Standrad Fail	RC/JU	ReRun
29	P4711-01	CF-613-COMP-16	BE101519.D	7 Nov 2024 5:25		RC/JU	Ok,M
30	P4711-06	CF-613-COMP-17	BE101520.D	7 Nov 2024 6:01	Need further 5X Dilution	RC/JU	Dilution
31	P4708-01	OR-02-110424	BE101521.D	7 Nov 2024 6:37		RC/JU	Ok,M
32	P4706-01	TR-04-110424	BE101522.D	7 Nov 2024 7:13		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110824

Review By	yogesh	Review On	11/10/2024 11:38:05 PM
Supervise By	mohammad	Supervise On	11/11/2024 12:24:16 AM
SubDirectory	BE110824	HP Acquire Method	BNA_E
		HP Processing Method	be110624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12326,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BE101543.D	8 Nov 2024 9:34		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BE101544.D	8 Nov 2024 10:07		RC/JU	Ok,M
3	PB164750BL	PB164750BL	BE101545.D	8 Nov 2024 10:51		RC/JU	Ok
4	PB164765BS	PB164765BS	BE101546.D	8 Nov 2024 11:27		RC/JU	Ok,M
5	PB164694TB	PB164694TB	BE101547.D	8 Nov 2024 12:03		RC/JU	Ok
6	P4701-08	BP-F4	BE101548.D	8 Nov 2024 12:45		RC/JU	Ok
7	P4700-04	MH-8	BE101549.D	8 Nov 2024 13:21		RC/JU	Ok
8	P4694-08	Z-04	BE101550.D	8 Nov 2024 13:56		RC/JU	Ok
9	P4694-04	Z-03A	BE101551.D	8 Nov 2024 14:32		RC/JU	Ok
10	P4738-02	72-12018	BE101552.D	8 Nov 2024 15:08		RC/JU	Ok
11	P4739-08	BP-G2	BE101553.D	8 Nov 2024 15:44		RC/JU	Ok
12	P4739-12	BP-B2	BE101554.D	8 Nov 2024 16:20		RC/JU	Ok
13	P4739-16	TP-11	BE101555.D	8 Nov 2024 16:55		RC/JU	Ok
14	P4739-04	TP-14	BE101556.D	8 Nov 2024 17:31		RC/JU	Ok
15	P4739-04MS	TP-14MS	BE101557.D	8 Nov 2024 18:07		RC/JU	Ok,M
16	P4739-04MSD	TP-14MSD	BE101558.D	8 Nov 2024 18:43		RC/JU	Ok,M
17	P4732-02	PPE-COMP	BE101559.D	8 Nov 2024 19:19		RC/JU	Ok
18	P4722-14	WC-3(0-6)	BE101560.D	8 Nov 2024 19:54		RC/JU	Ok

Instrument ID: BNA_E

Daily Analysis Runlog For Sequence/QC Batch ID # BE110824

Review By	yogesh	Review On	11/10/2024 11:38:05 PM
Supervise By	mohammad	Supervise On	11/11/2024 12:24:16 AM
SubDirectory	BE110824	HP Acquire Method	BNA_E
		HP Processing Method	be110624

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12326,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P4722-09	WC-2(0-6)	BE101561.D	8 Nov 2024 20:30		RC/JU	Ok
20	P4722-04	WC-1(0-6)	BE101562.D	8 Nov 2024 21:06		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110524

Review By	yogesh	Review On	11/6/2024 3:09:46 AM				
Supervise By	mohammad	Supervise On	11/6/2024 3:16:35 AM				
SubDirectory	BF110524	HP Acquire Method	BNA_F	HP Processing Method	bf110524		
STD. NAME		STD REF.#					
Tune/Reschk		SP6573					
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621					
CCC		SP6624					
Internal Standard/PEM		S12324,10ul/1000ul sample					
ICV/I.BLK		SP6559					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140230.D	05 Nov 2024 11:56		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF140231.D	05 Nov 2024 12:26		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF140232.D	05 Nov 2024 12:52	Compound#32,54,65 Removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF140233.D	05 Nov 2024 13:18	Compound#54 Kept on LR	RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF140234.D	05 Nov 2024 13:45		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF140235.D	05 Nov 2024 14:11	Calibration Good for DOD & 625.1 except for Benzidine	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF140236.D	05 Nov 2024 14:37	Com#77(Benzidine) Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF140237.D	05 Nov 2024 15:03		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF140238.D	05 Nov 2024 15:30		RC/JU	Ok,M
10	SSTDICV040	ICVBF110524	BF140239.D	05 Nov 2024 16:07		RC/JU	Ok
11	PB164366BL	PB164366BL	BF140240.D	05 Nov 2024 16:37		RC/JU	Ok
12	P4368-02	LOQ-SOIL-02-QT4-202	BF140241.D	05 Nov 2024 17:03	LOQ-SOIL-5 ppm	RC/JU	Ok
13	P4368-02	LOQ-SOIL-02-QT4-202	BF140242.D	05 Nov 2024 17:30	LOQ-SOIL-10 PPM (Used for All compounds except compounds#54,77)	RC/JU	Ok,M
14	P4368-08	LOQ-WATER-02-QT4-2	BF140243.D	05 Nov 2024 17:56	LOQ-Water-5 ppm	RC/JU	Ok,M
15	P4368-02	LOQ-SOIL-02-QT4-202	BF140244.D	05 Nov 2024 18:22	LOQ-SOIL-5 ppm, Already Analyzed	RC/JU	Not Ok
16	P4368-08	LOQ-WATER-02-QT4-2	BF140245.D	05 Nov 2024 18:49	LOQ-Water-10 ppm, Internal Standard Fail	RC/JU	Not Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110524

Review By	yogesh	Review On	11/6/2024 3:09:46 AM		
Supervise By	mohammad	Supervise On	11/6/2024 3:16:35 AM		
SubDirectory	BF110524	HP Acquire Method	BNA_F	HP Processing Method	bf110524
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12324,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

17	P4368-01	LOD-MDL-SOIL-01-QT	BF140246.D	05 Nov 2024 19:15	LOD-MDL-SOIL- 4 ppm	RC/JU	Ok
18	P4368-01	LOD-MDL-SOIL-01-QT	BF140247.D	05 Nov 2024 19:41	LOD-MDL-SOIL- 8 ppm	RC/JU	Ok,M
19	P4368-07	LOD-MDL-WATER-01-QT	BF140248.D	05 Nov 2024 20:07	LOD-MDL-WATER- 4 ppm	RC/JU	Ok,M
20	P4368-07	LOD-MDL-WATER-01-QT	BF140249.D	05 Nov 2024 20:34	LOD-MDL-WATER- 8 ppm	RC/JU	Ok,M
21	PB164369BL	PB164369BL	BF140250.D	05 Nov 2024 21:00		RC/JU	Ok
22	SSTDCCC040	SSTDCCC040EC	BF140251.D	05 Nov 2024 21:26		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110824

Review By	yogesh	Review On	11/10/2024 11:36:46 PM
Supervise By	mohammad	Supervise On	11/11/2024 12:24:00 AM
SubDirectory	BF110824	HP Acquire Method	BNA_F
		HP Processing Method	bf110524

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC	SP6624
Internal Standard/PEM	S12326,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140285.D	08 Nov 2024 09:09		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF140286.D	08 Nov 2024 09:35		RC/JU	Ok,M
3	PB164765BL	PB164765BL	BF140287.D	08 Nov 2024 10:02		RC/JU	Ok
4	PB164750BS	PB164750BS	BF140288.D	08 Nov 2024 10:28		RC/JU	Ok,M
5	PB164680BL	PB164680BL	BF140289.D	08 Nov 2024 10:55		RC/JU	Ok
6	PB164680BS	PB164680BS	BF140290.D	08 Nov 2024 11:22		RC/JU	Ok,M
7	P4732-01	PPE-COMP	BF140291.D	08 Nov 2024 11:54	Internal standard fail	RC/JU	Ok
8	P4694-05RE	Z-04RE	BF140292.D	08 Nov 2024 12:21	Internal standard fail	RC/JU	Confirms
9	P4720-01DL	JC-701-COMP-01DL	BF140293.D	08 Nov 2024 12:47		RC/JU	Ok
10	P4695-01DL	Z-01DL	BF140294.D	08 Nov 2024 13:13		RC/JU	Ok,M
11	P4739-01	TP-14	BF140295.D	08 Nov 2024 13:40	Internal standard fail	RC/JU	Ok
12	P4694-01DL	Z-03ADL	BF140296.D	08 Nov 2024 14:06	Internal standard fail	RC/JU	Ok,M
13	P4739-05	BP-G2	BF140297.D	08 Nov 2024 14:32	Internal standard fail	RC/JU	Ok
14	P4739-09	BP-B2	BF140298.D	08 Nov 2024 14:58	Internal Standard Fail, Need 5X Dilution	RC/JU	Dilution
15	P4737-01	TP2	BF140299.D	08 Nov 2024 15:24	Internal standard fail	RC/JU	Ok
16	P4737-01MS	TP2MS	BF140300.D	08 Nov 2024 15:51	Internal standard fail	RC/JU	Ok
17	P4737-01MSD	TP2MSD	BF140301.D	08 Nov 2024 16:17	Internal standard fail	RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF110824

Review By	yogesh	Review On	11/10/2024 11:36:46 PM		
Supervise By	mohammad	Supervise On	11/11/2024 12:24:00 AM		
SubDirectory	BF110824	HP Acquire Method	BNA_F	HP Processing Method	bf110524
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12326,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P4738-01	72-12018	BF140302.D	08 Nov 2024 16:43	Internal standard fail	RC/JU	ReRun
19	P4722-08	WC-2(0-6)	BF140303.D	08 Nov 2024 17:10	Internal standard fail, Need 2X Dilution	RC/JU	Dilution
20	P4739-13	TP-11	BF140304.D	08 Nov 2024 17:36	Internal Standard Fail	RC/JU	Ok,M
21	P4722-03	WC-1(0-6)	BF140305.D	08 Nov 2024 18:03	Internal Standard Fail	RC/JU	Ok,M
22	P4756-01	BP-B4	BF140306.D	08 Nov 2024 18:29	Internal Standard Fail	RC/JU	Ok
23	P4756-01MS	BP-B4MS	BF140307.D	08 Nov 2024 18:55	Internal Standard Fail	RC/JU	Ok
24	P4756-01MSD	BP-B4MSD	BF140308.D	08 Nov 2024 19:22	Internal Standard Fail	RC/JU	Ok
25	P4768-01	B-3-1	BF140309.D	08 Nov 2024 19:48	Internal Standard Fail	RC/JU	Ok
26	P4768-06	B-6-2	BF140310.D	08 Nov 2024 20:14	Internal standard fail, Need 2X Dilution	RC/JU	Dilution

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111124

Review By	yogesh	Review On	11/12/2024 1:20:41 AM
Supervise By	mohammad	Supervise On	11/13/2024 3:15:56 AM
SubDirectory	BF111124	HP Acquire Method	BNA_F
		HP Processing Method	bf111124

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6573 SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6624 S12326,10ul/1000ul sample SP6559

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF140311.D	11 Nov 2024 12:24		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF140312.D	11 Nov 2024 12:55		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF140313.D	11 Nov 2024 13:21	Compound #41,54 removed from 5ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF140314.D	11 Nov 2024 13:48		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF140315.D	11 Nov 2024 14:14		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF140316.D	11 Nov 2024 14:40	The Calibration is Good For 8270 DOD Except com#69,77 and good for 625.1 Method	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF140317.D	11 Nov 2024 15:07	Com#69,77 Failed in the calibration for both DOD and NON-DOD	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF140318.D	11 Nov 2024 15:34		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF140319.D	11 Nov 2024 16:00	Compound #9 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBF111124	BF140320.D	11 Nov 2024 16:30		RC/JU	Ok,M
11	PB164711BL	PB164711BL	BF140321.D	11 Nov 2024 17:01		RC/JU	Ok
12	PB164848BS	PB164848BS	BF140322.D	11 Nov 2024 17:32		RC/JU	Ok,M
13	PB164793TB	PB164793TB	BF140323.D	11 Nov 2024 17:59		RC/JU	Ok
14	P4771-06	P-1	BF140324.D	11 Nov 2024 18:25		RC/JU	Ok
15	P4771-08	P-2	BF140325.D	11 Nov 2024 18:52		RC/JU	Ok
16	P4788-04	BP-G3	BF140326.D	11 Nov 2024 19:18		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111124

Review By	yogesh	Review On	11/12/2024 1:20:41 AM		
Supervise By	mohammad	Supervise On	11/13/2024 3:15:56 AM		
SubDirectory	BF111124	HP Acquire Method	BNA_F	HP Processing Method	bf111124
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621			
CCC		SP6624			
Internal Standard/PEM		S12326,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

17	P4789-04	BP-F21	BF140327.D	11 Nov 2024 19:44		RC/JU	Ok
18	P4792-06	OILY-STONE-DRUM	BF140328.D	11 Nov 2024 20:10		RC/JU	Ok
19	P4798-04	MH-6	BF140329.D	11 Nov 2024 20:36		RC/JU	Ok
20	P4718-03	WB-307-SB02	BF140330.D	11 Nov 2024 21:03		RC/JU	Ok

M : Manual Integration