

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102824

Review By	Mahesh Dadoda	Review On	10/29/2024 12:19:57 PM
Supervise By	Semsettin Yesilyurt	Supervise On	10/29/2024 12:21:34 PM
SubDirectory	VX102824	HP Acquire Method	MSVOA_X HP Processing Method 82x102824w.m
STD. NAME	STD REF.#		
Tune/Reschk	VP131142		
Initial Calibration Stds	VP131143,VP131144,VP131145,VP131147,VP131148,VP131149		
CCC	VP131151,VP131152		
Internal Standard/PEM	VP128298		
ICV/I.BLK	VP131150		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX043555.D	28 Oct 2024 10:09		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX043556.D	28 Oct 2024 10:59	Method pass	JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX043557.D	28 Oct 2024 11:22	8260D	JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX043558.D	28 Oct 2024 11:45		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX043559.D	28 Oct 2024 12:08		JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX043560.D	28 Oct 2024 12:31		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX043561.D	28 Oct 2024 12:54		JC/MD	Ok,M
8	VIBLK	VIBLK	VX043562.D	28 Oct 2024 14:04		JC/MD	Ok
9	VSTDICV050	ICVVX102824	VX043563.D	28 Oct 2024 14:27		JC/MD	Ok,M
10	VX1028MBL01	VX1028MBL01	VX043564.D	28 Oct 2024 15:02	Internal standard fail	JC/MD	Not Ok
11	VX1028WBL01	VX1028WBL01	VX043565.D	28 Oct 2024 15:52		JC/MD	Ok
12	VX1028WBS01	VX1028WBS01	VX043566.D	28 Oct 2024 16:28		JC/MD	Ok,M
13	VX1028WBSD01	VX1028WBSD01	VX043567.D	28 Oct 2024 16:51		JC/MD	Ok,M
14	PB164394TB	PB164394TB	VX043568.D	28 Oct 2024 17:14	pH#5.0 A	JC/MD	Ok
15	PB164394ZHE#01	PB164394ZHE#01	VX043569.D	28 Oct 2024 17:38	pH#5.0 A	JC/MD	Ok
16	PB164394ZHE#02	PB164394ZHE#02	VX043570.D	28 Oct 2024 18:01	pH#5.0 A	JC/MD	Ok
17	PB164394ZHE#03	PB164394ZHE#03	VX043571.D	28 Oct 2024 18:24	pH#5.0 A	JC/MD	Ok
18	PB164394ZHE#04	PB164394ZHE#04	VX043572.D	28 Oct 2024 18:47	pH#5.0 A	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX102824

Review By	Mahesh Dadoda	Review On	10/29/2024 12:19:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	10/29/2024 12:21:34 PM		
SubDirectory	VX102824	HP Acquire Method	MSVOA_X	HP Processing Method	82x102824w.m
STD. NAME		STD REF.#			
Tune/Reschk		VP131142			
Initial Calibration Stds		VP131143,VP131144,VP131145,VP131147,VP131148,VP131149			
CCC		VP131151,VP131152			
Internal Standard/PEM		VP128298			
ICV/I.BLK		VP131150			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	PB164394ZHE#05	PB164394ZHE#05	VX043573.D	28 Oct 2024 19:10	pH#5.0 A	JC/MD	Ok
20	PB164394ZHE#06	PB164394ZHE#06	VX043574.D	28 Oct 2024 19:33	pH#5.0 A	JC/MD	Ok
21	PB164394ZHE#07	PB164394ZHE#07	VX043575.D	28 Oct 2024 19:56	pH#5.0 A	JC/MD	Ok
22	PB164394ZHE#08	PB164394ZHE#08	VX043576.D	28 Oct 2024 20:19	pH#5.0 A	JC/MD	Ok
23	PB164394ZHE#09	PB164394ZHE#09	VX043577.D	28 Oct 2024 20:42	pH#5.0 A	JC/MD	Ok
24	PB164394ZHE#10	PB164394ZHE#10	VX043578.D	28 Oct 2024 21:05	pH#5.0 A	JC/MD	Ok
25	PB164394ZHE#11	PB164394ZHE#11	VX043579.D	28 Oct 2024 21:28	pH#5.0 A	JC/MD	Ok
26	PB164394ZHE#12	PB164394ZHE#12	VX043580.D	28 Oct 2024 21:52	pH#5.0 A	JC/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VX043581.D	28 Oct 2024 22:15	out of tune time	JC/MD	Not Ok

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX110724

Review By	Mahesh Dadoda	Review On	11/8/2024 2:39:29 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	11/8/2024 2:42:03 PM		
SubDirectory	VX110724	HP Acquire Method	MSVOA_X	HP Processing Method	82X102824W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131334
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131337,VP131338

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX043725.D	07 Nov 2024 10:18		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX043726.D	07 Nov 2024 10:46		JC/MD	Ok,M
3	VX1107MBL01	VX1107MBL01	VX043727.D	07 Nov 2024 11:29		JC/MD	Ok
4	VX1107WBL01	VX1107WBL01	VX043728.D	07 Nov 2024 11:53		JC/MD	Ok
5	VX1107WBS01	VX1107WBS01	VX043729.D	07 Nov 2024 12:19		JC/MD	Ok,M
6	VX1107WBSD01	VX1107WBSD01	VX043730.D	07 Nov 2024 12:42	BSD high for com #40	JC/MD	Ok,M
7	VIBLK	VIBLK	VX043731.D	07 Nov 2024 13:10		JC/MD	Ok
8	PB164695TB	PB164695TB	VX043732.D	07 Nov 2024 13:33	Surrogate fail,pH#5.0 A	JC/MD	ReRun
9	P4718-03	WB-307-SB02	VX043733.D	07 Nov 2024 13:56	,pH#6.0 A	JC/MD	Ok
10	P4719-03	BAYAVE-STOCKPILE	VX043734.D	07 Nov 2024 14:19	Surrogate fail,pH#5.0 A	JC/MD	ReRun
11	P4720-05	JC-701-COMP-01	VX043735.D	07 Nov 2024 14:43	Surrogate fail,pH#5.0 A	JC/MD	ReRun
12	P4722-02	WC-1(3.5)	VX043736.D	07 Nov 2024 15:06	Surrogate fail,pH#7.0 A	JC/MD	ReRun
13	P4722-07	WC-2(4)	VX043737.D	07 Nov 2024 15:29	Surrogate fail,pH#5.0 A	JC/MD	ReRun
14	P4722-12	WC-3(3)	VX043738.D	07 Nov 2024 15:52	Surrogate fail,pH#7.0 A	JC/MD	ReRun
15	P4722-16	WC-1(5.5)	VX043739.D	07 Nov 2024 16:16	Surrogate fail,pH#7.0 A	JC/MD	ReRun
16	P4722-18	WC-2(2)	VX043740.D	07 Nov 2024 16:39	Surrogate fail,pH#7.0 A	JC/MD	ReRun
17	P4722-20	WC-3(5)	VX043741.D	07 Nov 2024 17:02	Surrogate fail,pH#6.0 A	JC/MD	ReRun
18	P4732-05	PPE-Grab	VX043742.D	07 Nov 2024 17:25	Surrogate fail,pH#5.0 A	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX110724

Review By	Mahesh Dadoda	Review On	11/8/2024 2:39:29 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	11/8/2024 2:42:03 PM		
SubDirectory	VX110724	HP Acquire Method	MSVOA_X	HP Processing Method	82X102824W.M
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP131334			
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP131337,VP131338			

19	P4738-02	72-12018	VX043743.D	07 Nov 2024 17:49	Surrogate fail,pH#5.0 A	JC/MD	ReRun
20	P4739-04	TP-14	VX043744.D	07 Nov 2024 18:12	Surrogate fail,pH#5.0 A	JC/MD	ReRun
21	P4739-08	BP-G2	VX043745.D	07 Nov 2024 18:35	BSD high; Surrogate fail,pH#5.0 A	JC/MD	ReRun
22	P4739-12	BP-B2	VX043746.D	07 Nov 2024 18:58	BSD high; Surrogate fail,pH#5.0 A	JC/MD	ReRun
23	P4739-16	TP-11	VX043747.D	07 Nov 2024 19:21	,pH#7.0 A	JC/MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VX043748.D	07 Nov 2024 19:45		JC/MD	Ok,M

M : Manual Integration