

Surrogate Summary

SDG No.: P4718

Client: Portal Partners Tri-Venture

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4718-03	WB-307-SB02	1,2-Dichloroethane-d4	50	36.8	74	70 (74)	130 (125)
		Dibromofluoromethane	50	43.7	87	70 (75)	130 (124)
		Toluene-d8	50	49.7	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.0	90	70 (77)	130 (121)

Surrogate Summary

SDG No.: P4718

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX1107WBL01	VX1107WBL01	1,2-Dichloroethane-d4	50	38.1	76	70 (74)	130 (125)
		Dibromofluoromethane	50	45.9	92	70 (75)	130 (124)
		Toluene-d8	50	48.8	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.7	93	70 (77)	130 (121)
VX1107WBS01	VX1107WBS01	1,2-Dichloroethane-d4	50	42.8	86	70 (74)	130 (125)
		Dibromofluoromethane	50	47.6	95	70 (75)	130 (124)
		Toluene-d8	50	48.7	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.8	98	70 (77)	130 (121)
VX1107WBSD0	VX1107WBSD01	1,2-Dichloroethane-d4	50	43.0	86	70 (74)	130 (125)
		Dibromofluoromethane	50	48.5	97	70 (75)	130 (124)
		Toluene-d8	50	47.8	96	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.6	95	70 (77)	130 (121)

