

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4718

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Datafile : VX043729.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1107WBS01	Vinyl chloride	20	17.9	ug/L	90			70 (65)	130 (117)	
	1,1-Dichloroethene	20	20.2	ug/L	101			70 (74)	130 (110)	
	2-Butanone	100	100	ug/L	100			40 (65)	160 (122)	
	Carbon Tetrachloride	20	18.8	ug/L	94			70 (77)	130 (113)	
	Chloroform	20	22.4	ug/L	112			70 (79)	130 (113)	
	Benzene	20	21.5	ug/L	108			70 (82)	130 (109)	
	1,2-Dichloroethane	20	20.0	ug/L	100			70 (80)	130 (115)	
	Trichloroethene	20	21.9	ug/L	110			70 (77)	130 (113)	
	Tetrachloroethene	20	20.5	ug/L	103			70 (67)	130 (123)	
	Chlorobenzene	20	22.1	ug/L	111			70 (82)	130 (109)	