

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718
 Lab File ID: VX043726.D Date Analyzed: 11/07/2024
 Instrument ID: MSVOA_X Time Analyzed: 10:46
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	96887	5.54	175036	6.76	158336	10.05
UPPER LIMIT	193774	6.044	350072	7.257	316672	10.549
LOWER LIMIT	48443.5	5.044	87518	6.257	79168	9.549
EPA SAMPLE NO.						
WB-307-SB02	74982	5.55	134619	6.76	118244	10.06
VX1107WBL01	93207	5.55	162976	6.76	144247	10.06
VX1107WBS01	82126	5.54	153670	6.76	142458	10.06
VX1107WBSD01	83388	5.55	159680	6.76	147769	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.