

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718  
 Lab File ID: VX043726.D Date Analyzed: 11/07/2024  
 Instrument ID: MSVOA\_X Time Analyzed: 10:46  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 77959         | 12.018 |  |  |  |  |
| UPPER LIMIT    | 155918        | 12.518 |  |  |  |  |
| LOWER LIMIT    | 38979.5       | 11.518 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| WB-307-SB02    | 50308         | 12.02  |  |  |  |  |
| VX1107WBL01    | 66166         | 12.02  |  |  |  |  |
| VX1107WBS01    | 70080         | 12.02  |  |  |  |  |
| VX1107WBSD01   | 69442         | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.