

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG No.: P4718

Instrument ID: MSVOA_Y Calibration Date/Time: 11/08/2024 09:33

Lab File ID: VY020221.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:39 14:52

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.463	0.412		-11.02	20
Chloromethane	0.767	0.758	0.1	-1.17	20
Vinyl Chloride	0.823	0.902		9.6	20
Bromomethane	0.546	0.575		5.31	20
Chloroethane	0.531	0.600		12.99	20
Trichlorofluoromethane	1.005	1.166		16.02	20
1,1,2-Trichlorotrifluoroethane	0.564	0.638		13.12	20
1,1-Dichloroethene	0.540	0.596		10.37	20
Acetone	0.145	0.177		22.07	20
Carbon Disulfide	1.796	1.901		5.85	20
Methyl tert-butyl Ether	1.357	1.510		11.27	20
Methyl Acetate	0.340	0.345		1.47	20
Methylene Chloride	0.652	0.630		-3.37	20
trans-1,2-Dichloroethene	0.609	0.682		11.99	20
1,1-Dichloroethane	1.170	1.333	0.1	13.93	20
Cyclohexane	1.132	1.185		4.68	20
2-Butanone	0.197	0.214		8.63	20
Carbon Tetrachloride	0.494	0.582		17.81	20
cis-1,2-Dichloroethene	0.700	0.800		14.29	20
Bromochloromethane	0.548	0.552		0.73	20
Chloroform	1.161	1.349		16.19	20
1,1,1-Trichloroethane	1.013	1.203		18.76	20
Methylcyclohexane	0.623	0.708		13.64	20
Benzene	1.433	1.608		12.21	20
1,2-Dichloroethane	0.397	0.432		8.82	20
Trichloroethene	0.332	0.376		13.25	20
1,2-Dichloropropane	0.343	0.385		12.24	20
Bromodichloromethane	0.487	0.557		14.37	20
4-Methyl-2-Pentanone	0.232	0.242		4.31	20
Toluene	0.884	1.030		16.52	20
t-1,3-Dichloropropene	0.430	0.481		11.86	20
cis-1,3-Dichloropropene	0.513	0.573		11.7	20
1,1,2-Trichloroethane	0.231	0.255		10.39	20
2-Hexanone	0.166	0.179		7.83	20
Dibromochloromethane	0.295	0.334		13.22	20
1,2-Dibromoethane	0.215	0.231		7.44	20
Tetrachloroethene	0.330	0.386		16.97	20
Chlorobenzene	1.065	1.227	0.3	15.21	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG No.: P4718

Instrument ID: MSVOA_Y Calibration Date/Time: 11/08/2024 09:33

Lab File ID: VY020221.D Init. Calib. Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:39 14:52

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.992	2.350		17.97	20
m/p-Xylenes	0.729	0.859		17.83	20
o-Xylene	0.690	0.808		17.1	20
Styrene	1.141	1.371		20.16	20
Bromoform	0.175	0.201	0.1	14.86	20
Isopropylbenzene	3.994	4.721		18.2	20
1,1,2,2-Tetrachloroethane	0.699	0.726	0.3	3.86	20
1,3-Dichlorobenzene	1.641	1.877		14.38	20
1,4-Dichlorobenzene	1.634	1.830		11.99	20
1,2-Dichlorobenzene	1.426	1.635		14.66	20
1,2-Dibromo-3-Chloropropane	0.106	0.107		0.94	20
1,2,4-Trichlorobenzene	0.766	0.872		13.84	20
1,2,3-Trichlorobenzene	0.650	0.716		10.15	20
1,2-Dichloroethane-d4	0.624	0.603		-3.37	20
Dibromofluoromethane	0.318	0.309		-2.83	20
Toluene-d8	1.265	1.242		-1.82	20
4-Bromofluorobenzene	0.411	0.414		0.73	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.