

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718
 Lab File ID: VN084842.D Date Analyzed: 11/14/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:54
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	192160	8.22	316615	9.09	272995	11.87
UPPER LIMIT	384320	8.718	633230	9.594	545990	12.365
LOWER LIMIT	96080	7.718	158308	8.594	136498	11.365
EPA SAMPLE NO.						
WB-307-SW	203676	8.22	345476	9.09	299825	11.87
TB-11042024	193262	8.22	335586	9.10	288886	11.87
VN1114WBL01	203029	8.22	340637	9.09	293869	11.86
VN1114WBS02	169550	8.22	270058	9.09	234639	11.87
VN1114WBSD02	161141	8.22	263729	9.10	229436	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.