

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718
 Lab File ID: VN084842.D Date Analyzed: 11/14/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:54
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	137639	13.788				
UPPER LIMIT	275278	14.288				
LOWER LIMIT	68819.5	13.288				
EPA SAMPLE NO.						
WB-307-SW	136300	13.79				
TB-11042024	136820	13.79				
VN1114WBL01	131669	13.79				
VN1114WBS02	118912	13.79				
VN1114WBSD02	112382	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.