

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718
 Lab File ID: VY020221.D Date Analyzed: 11/08/2024
 Instrument ID: MSVOA_Y Time Analyzed: 09:33
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	208354	13.353				
UPPER LIMIT	416708	13.853				
LOWER LIMIT	104177	12.853				
EPA SAMPLE NO.						
WB-307-SB01	128433	13.35				
WB-307-SB02	136217	13.35				
VY1108SBL01	137264	13.35				
VY1108SBS01	187903	13.35				
VY1108SBSD01	193395	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.