

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: P4718 SAS No.: P4718 SDG NO.: P4718
Lab File ID: VY020074.D BFB Injection Date: 10/30/2024
Instrument ID: MSVOA_Y BFB Injection Time: 12:07
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	24
75	30.0 - 60.0% of mass 95	57.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	1.1 (1.4) 1
174	50.0 - 100.0% of mass 95	74.5
175	5.0 - 9.0% of mass 174	5.8 (7.8) 1
176	95.0 - 101.0% of mass 174	70.8 (95.1) 1
177	5.0 - 9.0% of mass 176	4.9 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY020075.D	10/30/2024	12:39
VSTDIC010	VSTDIC010	VY020076.D	10/30/2024	13:02
VSTDIC020	VSTDIC020	VY020077.D	10/30/2024	13:24
VSTDIC050	VSTDIC050	VY020078.D	10/30/2024	14:06
VSTDIC100	VSTDIC100	VY020079.D	10/30/2024	14:29
VSTDIC150	VSTDIC150	VY020080.D	10/30/2024	14:52