

CASE NARRATIVE

ENTACT

Project Name: North Point

Project # N/A

Chemtech Project # P4721

Test Name: EPH

A. Number of Samples and Date of Receipt:

15 Solid samples were received on 11/05/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Diesel Range Organics, EPH and Gasoline Range Organics. This data package contains results for EPH.

C. Analytical Techniques:

The analysis were performed on instrument FID_C. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analyses were performed on instrument FID_D. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analysis were performed on instrument FID_E. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 10224. The analysis were performed on instrument FID_G. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 13302. The analysis were performed on instrument FID_F. The column is RXI-1MS which is 20 meters, 0.18mm ID, 0.18 um df, catalog 13302. The analysis of EPHs was based on method NJEPH and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for

EX-2-TPH-1DL [2-Bromonaphthalene (SURR) - 180%],
EX-2-TPH-2DL [2-Bromonaphthalene (SURR) - 155%],
EX-2-TPH-3 [2-Bromonaphthalene (SURR) - 140%],
EX-2-TPH-3DL [2-Bromonaphthalene (SURR) - 160%],
EX-11-TPH-1 [2-Bromonaphthalene (SURR) - 149%],
EX-11-TPH-1DL [2-Bromonaphthalene (SURR) - 170%, 2-Fluorobiphenyl (SURR) - 146%],
EX-11-TPH-1DL2 [2-Bromonaphthalene (SURR) - 174%],
EX-11-TPH-2 [2-Bromonaphthalene (SURR) - 150%],
EX-11-TPH-2DL [2-Bromonaphthalene (SURR) - 181%],
EX-11-TPH-4 [2-Bromonaphthalene (SURR) - 163%, 2-Fluorobiphenyl (SURR) - 147%],
EX-11-TPH-4DL [2-Bromonaphthalene (SURR) - 183%, 2-Fluorobiphenyl (SURR) - 165%],
EX-11-TPH-4DL2 [2-Bromonaphthalene (SURR) - 208%, 2-Fluorobiphenyl

(SURRE) - 188%], EX-9-TPH-7DL2 [2-Bromonaphthalene (SURRE) - 168% and 2-Fluorobiphenyl (SURRE) - 165%] due to high concentration sample required dilution therefore samples were reanalyzed with dilution and reported

while, EX-2-TPH-1DL2 [1-chlorooctadecane (SURRE) - 0%], EX-11-TPH-2DL [1-chlorooctadecane (SURRE) - 0%], EX-11-TPH-4DL2 [1-chlorooctadecane (SURRE) - 0%], Surrogates were diluted out due to the high dilution No further corrective action was taken.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds .

The MSD recoveries met the acceptable requirements .

The RPD for {P4718-01MSD} with File ID: FD048679.D met criteria except for Aromatic C12-C16[28.9%] due to difference in results of MS-MSD.

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

Sample EX-11-TPH-2 was diluted due to extremely bad matrix.

Samples EX-2-TPH-1, EX-2-TPH-2, EX-2-TPH-3, EX-11-TPH-1, EX-11-TPH-2, EX-11-TPH-3, EX-11-TPH-4, EX-9-TPH-4 and EX-9-TPH-7 EX-2-TPH-1DL, EX-2-TPH-2DL, EX-11-TPH-1DL, EX-11-TPH-4DL, EX-9-TPH-7DL were diluted due to high concentrations.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922