

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: ENTA05
 ProjectID: North Point
 Lab Code: CHEM Case No.: P4721 SAS No.: P4721 SDG No.: P4721

Calibration Sequence : FG102124		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	124758474	124758	FG014453.D
500	63252619	126505	FG014454.D
200	26227964	131140	FG014455.D
100	13573191	135732	FG014456.D
50	6864383	137288	FG014457.D
AVG RF : 131085		% RSD : 4.197	AVG RT : 15.0822