

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: ENTA05
ProjectID: North Point
Lab Code: CHEM Case No.: P4721 SAS No.: P4721 SDG No.: P4721
DataFile: FG014669.D Analyst Name: YP\AJ Analyst Date: 11-07-2024

Conc. (PPM)	Area Count	RF	Average RF	%D
500	65518705	131037	131085	0.037