



SAMPLE DATA

Report of Analysis

Client:	ENTACT	Date Collected:	11/05/24
Project:	North Point	Date Received:	11/05/24
Client Sample ID:	EX-2-TPH-1	SDG No.:	P4721
Lab Sample ID:	P4721-01	Matrix:	SOIL
Analytical Method:	8015D DRO	% Solid:	86.7
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Diesel Range Organics
GPC Factor :		Injection Volume :	
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014694.D	100	11/07/24 08:30	11/08/24 3:16	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	1360000		21300	192000	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	0.00	*	37 - 130	0%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	11/05/24	
Project:	North Point		Date Received:	11/05/24	
Client Sample ID:	EX-2-TPH-2		SDG No.:	P4721	
Lab Sample ID:	P4721-02		Matrix:	SOIL	
Analytical Method:	8015D DRO		% Solid:	83	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	Diesel Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014695.D	100	11/07/24 08:30	11/08/24 3:44	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	1270000		22200	200000	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	0.00	*	37 - 130	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	11/05/24	
Project:	North Point		Date Received:	11/05/24	
Client Sample ID:	EX-2-TPH-3		SDG No.:	P4721	
Lab Sample ID:	P4721-03		Matrix:	SOIL	
Analytical Method:	8015D DRO		% Solid:	88.3	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	Diesel Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014696.D	50	11/07/24 08:30	11/08/24 4:12	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	430000		10400	94200	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	0.00	*	37 - 130	0%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	11/05/24	
Project:	North Point		Date Received:	11/05/24	
Client Sample ID:	EX-11-TPH-1		SDG No.:	P4721	
Lab Sample ID:	P4721-04		Matrix:	SOIL	
Analytical Method:	8015D DRO		% Solid:	79.7	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	Diesel Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014697.D	250	11/07/24 08:30	11/08/24 4:41	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	2830000		57900	522000	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	0.00	*	37 - 130	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT	Date Collected:	11/05/24
Project:	North Point	Date Received:	11/05/24
Client Sample ID:	EX-11-TPH-2	SDG No.:	P4721
Lab Sample ID:	P4721-05	Matrix:	SOIL
Analytical Method:	8015D DRO	% Solid:	91.5
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Diesel Range Organics
GPC Factor :		Injection Volume :	
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014698.D	2500	11/07/24 08:30	11/08/24 5:09	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	12100000		505000	4550000	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	0.00	*	37 - 130	0%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT	Date Collected:	11/05/24
Project:	North Point	Date Received:	11/05/24
Client Sample ID:	EX-11-TPH-3	SDG No.:	P4721
Lab Sample ID:	P4721-06	Matrix:	SOIL
Analytical Method:	8015D DRO	% Solid:	79.4
Sample Wt/Vol:	30.1	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Diesel Range Organics
GPC Factor :		Injection Volume :	
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014676.D	10	11/07/24 08:30	11/07/24 15:56	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	58200		2320	20900	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	1.17		37 - 130	58%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	11/05/24	
Project:	North Point		Date Received:	11/05/24	
Client Sample ID:	EX-11-TPH-4		SDG No.:	P4721	
Lab Sample ID:	P4721-07		Matrix:	SOIL	
Analytical Method:	8015D DRO		% Solid:	85.6	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	Diesel Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014706.D	100	11/07/24 08:30	11/08/24 10:24	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	1590000		21600	195000	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	0.00	*	37 - 130	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	11/05/24	
Project:	North Point		Date Received:	11/05/24	
Client Sample ID:	EX-9-TPH-1		SDG No.:	P4721	
Lab Sample ID:	P4721-08		Matrix:	SOIL	
Analytical Method:	8015D DRO		% Solid:	83.7	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	Diesel Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014702.D	1	11/07/24 08:30	11/08/24 8:31	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	2680		220	1990	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	9.89		37 - 130	49%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	11/05/24	
Project:	North Point		Date Received:	11/05/24	
Client Sample ID:	EX-9-TPH-2		SDG No.:	P4721	
Lab Sample ID:	P4721-09		Matrix:	SOIL	
Analytical Method:	8015D DRO		% Solid:	85.5	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	Diesel Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014683.D	1	11/07/24 08:30	11/07/24 20:39	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	4790		216	1950	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	14.5		37 - 130	72%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT	Date Collected:	11/05/24
Project:	North Point	Date Received:	11/05/24
Client Sample ID:	EX-9-TPH-3	SDG No.:	P4721
Lab Sample ID:	P4721-10	Matrix:	SOIL
Analytical Method:	8015D DRO	% Solid:	85.2
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Final Vol:	1
GPC Factor :		PH :	
Prep Method :	SW3541	Decanted:	
		Test:	Diesel Range Organics
		Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014686.D	10	11/07/24 08:30	11/07/24 22:05	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	31600		2170	19500	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	1.30		37 - 130	65%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	11/05/24	
Project:	North Point		Date Received:	11/05/24	
Client Sample ID:	EX-9-TPH-4		SDG No.:	P4721	
Lab Sample ID:	P4721-11		Matrix:	SOIL	
Analytical Method:	8015D DRO		% Solid:	89.4	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	Diesel Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014687.D	10	11/07/24 08:30	11/07/24 22:33	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	229000		2060	18600	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	2.04		37 - 130	102%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	11/05/24	
Project:	North Point		Date Received:	11/05/24	
Client Sample ID:	EX-9-TPH-5		SDG No.:	P4721	
Lab Sample ID:	P4721-12		Matrix:	SOIL	
Analytical Method:	8015D DRO		% Solid:	86.9	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	Diesel Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014703.D	1	11/07/24 08:30	11/08/24 8:59	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	2400		212	1920	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	19.0		37 - 130	95%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	11/05/24	
Project:	North Point		Date Received:	11/05/24	
Client Sample ID:	EX-9-TPH-6		SDG No.:	P4721	
Lab Sample ID:	P4721-13		Matrix:	SOIL	
Analytical Method:	8015D DRO		% Solid:	83.6	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	1	mL
Soil Aliquot Vol:		uL	Test:	Diesel Range Organics	
Extraction Type:			Injection Volume :		
GPC Factor :		PH :			
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014704.D	1	11/07/24 08:30	11/08/24 9:27	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	1950	J	221	1990	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	11.4		37 - 130	57%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT	Date Collected:	11/05/24
Project:	North Point	Date Received:	11/05/24
Client Sample ID:	EX-9-TPH-7	SDG No.:	P4721
Lab Sample ID:	P4721-14	Matrix:	SOIL
Analytical Method:	8015D DRO	% Solid:	86.8
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Diesel Range Organics
GPC Factor :		Injection Volume :	
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014707.D	200	11/07/24 08:30	11/08/24 10:52	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	2970000		42600	384000	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	0.00	*	37 - 130	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT	Date Collected:	11/05/24
Project:	North Point	Date Received:	11/05/24
Client Sample ID:	EX-9-TPH-8	SDG No.:	P4721
Lab Sample ID:	P4721-15	Matrix:	SOIL
Analytical Method:	8015D DRO	% Solid:	87.2
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Diesel Range Organics
GPC Factor :		Injection Volume :	
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
FG014705.D	1	11/07/24 08:30	11/08/24 9:56	PB164744

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
DRO	DRO	1760	J	212	1910	ug/kg
SURROGATES						
16416-32-3	Tetracosane-d50	9.09		37 - 130	45%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	P4721	OrderDate:	11/5/2024 3:27:00 PM
Client:	ENTACT	Project:	North Point
Contact:	Wyatt Seel	Location:	L23

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P4721-01	EX-2-TPH-1	SOIL	Diesel Range Organics	8015D	11/05/24	11/07/24	11/08/24	11/05/24
P4721-02	EX-2-TPH-2	SOIL	Diesel Range Organics	8015D	11/05/24	11/07/24	11/08/24	11/05/24
P4721-03	EX-2-TPH-3	SOIL	Diesel Range Organics	8015D	11/05/24	11/07/24	11/08/24	11/05/24
P4721-04	EX-11-TPH-1	SOIL	Diesel Range Organics	8015D	11/05/24	11/07/24	11/08/24	11/05/24
P4721-05	EX-11-TPH-2	SOIL	Diesel Range Organics	8015D	11/05/24	11/07/24	11/08/24	11/05/24
P4721-06	EX-11-TPH-3	SOIL	Diesel Range Organics	8015D	11/05/24	11/07/24	11/07/24	11/05/24
P4721-07	EX-11-TPH-4	SOIL	Diesel Range Organics Gasoline Range Organics	8015D 8015D	11/05/24	11/07/24	11/08/24 11/07/24	11/05/24
P4721-08	EX-9-TPH-1	SOIL	Diesel Range Organics	8015D	11/05/24	11/07/24	11/08/24	11/05/24
P4721-09	EX-9-TPH-2	SOIL	Diesel Range Organics	8015D	11/05/24	11/07/24	11/07/24	11/05/24
P4721-10	EX-9-TPH-3	SOIL	Diesel Range Organics	8015D	11/05/24	11/07/24	11/07/24	11/05/24
P4721-11	EX-9-TPH-4	SOIL	Diesel Range Organics	8015D	11/05/24	11/07/24	11/07/24	11/05/24

LAB CHRONICLE

P4721-12	EX-9-TPH-5	SOIL			11/05/24			11/05/24
			Diesel Range Organics	8015D		11/07/24	11/08/24	
			Gasoline Range Organics	8015D			11/07/24	
P4721-13	EX-9-TPH-6	SOIL			11/05/24			11/05/24
			Diesel Range Organics	8015D		11/07/24	11/08/24	
P4721-14	EX-9-TPH-7	SOIL			11/05/24			11/05/24
			Diesel Range Organics	8015D		11/07/24	11/08/24	
P4721-15	EX-9-TPH-8	SOIL			11/05/24			11/05/24
			Diesel Range Organics	8015D		11/07/24	11/08/24	