

SOIL DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RI

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4721 **SAS No :** P4721 **SDG No:** P4721
Matrix Spike - EPA Sample No : PB164744BS **Datafile:** FG014682.D

COMPOUND	SPIKE ADDED ug/kg	CONCENTRATION ug/kg	LCS/LCSD CONCENTRATION ug/kg	% REC	QC LIMITS
DRO	6660	0	7371	111	68-131