

Continuing Calibration Report for SequenceID : FC110824AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC067683.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 7306679.000  | 60.000  | 3.171  | 6.450  | 121777.983 | 135469.641 | 10.107 |
| Aliphatic C12-C16 | 4818056.000  | 40.000  | 6.451  | 9.841  | 120451.400 | 131681.730 | 8.528  |
| Aliphatic C16-C21 | 7034927.000  | 60.000  | 9.842  | 13.199 | 117248.783 | 127235.092 | 7.849  |
| Aliphatic C21-C28 | 9100518.000  | 80.000  | 13.200 | 16.853 | 113756.475 | 120090.398 | 5.274  |
| Aliphatic C28-C40 | 12781161.000 | 120.000 | 16.854 | 21.698 | 106509.675 | 101093.195 | -5.358 |
| Aliphatic EPH     | 41041341.000 | 360.000 | 3.171  | 21.698 | 114003.725 | 118799.912 | 4.037  |