

Continuing Calibration Report for SequenceID : FC111124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067686.D

Aliphatic C9-C12	8560870.000	60.000	3.170	6.450	142681.167	135469.641	-5.323
Aliphatic C12-C16	5589068.000	40.000	6.451	9.840	139726.700	131681.730	-6.109
Aliphatic C16-C21	8067406.000	60.000	9.841	13.198	134456.767	127235.092	-5.676
Aliphatic C21-C28	10134643.000	80.000	13.199	16.853	126683.038	120090.398	-5.490
Aliphatic C28-C40	12758319.000	120.000	16.854	21.695	106319.325	101093.195	-5.170
Aliphatic EPH	45110306.000	360.000	3.170	21.695	125306.406	118799.912	-5.477