

Continuing Calibration Report for SequenceID : FC111124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067693.D

Aliphatic C9-C12	8271307.000	60.000	3.170	6.450	137855.117	135469.641	-1.761
Aliphatic C12-C16	5396973.000	40.000	6.451	9.840	134924.325	131681.730	-2.462
Aliphatic C16-C21	7819384.000	60.000	9.841	13.198	130323.067	127235.092	-2.427
Aliphatic C21-C28	9814622.000	80.000	13.199	16.853	122682.775	120090.398	-2.159
Aliphatic C28-C40	12463555.000	120.000	16.854	21.695	103862.958	101093.195	-2.740
Aliphatic EPH	43765841.000	360.000	3.170	21.695	121571.781	118799.912	-2.333