

Continuing Calibration Report for SequenceID : FD110824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048672.D

Aromatic C10-C12	7919867.000	40.000	4.092	5.810	197996.675	176202.694	-12.369
Aromatic C12-C16	11889358.000	60.000	5.811	8.415	198155.967	177273.493	-11.780
Aromatic C16-C21	13817573.000	80.000	8.416	12.678	172719.663	163832.550	-5.425
Aromatic C21-C36	31272382.000	180.000	12.679	18.083	173735.456	152713.153	-13.766
Aromatic EPH	64899180.000	360.000	4.092	18.083	180275.500	161887.469	-11.359