

Continuing Calibration Report for SequenceID : FD111124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048689.D

Aromatic C10-C12	7120820.000	40.000	4.090	5.808	178020.500	176202.694	-1.032
Aromatic C12-C16	10580549.000	60.000	5.809	8.414	176342.483	177273.493	0.525
Aromatic C16-C21	12679055.000	80.000	8.415	12.677	158488.188	163832.550	3.262
Aromatic C21-C36	27496895.000	180.000	12.678	18.084	152760.528	152713.153	-0.031
Aromatic EPH	57877319.000	360.000	4.090	18.084	160770.331	161887.469	0.690