

Continuing Calibration Report for SequenceID : FE110924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE051134.D

Aliphatic C9-C12	9357286.000	60.000	3.140	6.758	155954.767	139914.732	-11.464
Aliphatic C12-C16	6119922.000	40.000	6.759	10.191	152998.050	140635.568	-8.790
Aliphatic C16-C21	8890102.000	60.000	10.192	13.550	148168.367	137768.027	-7.549
Aliphatic C21-C28	11654865.000	80.000	13.551	17.206	145685.813	133951.574	-8.760
Aliphatic C28-C40	15222579.000	120.000	17.207	22.053	126854.825	128748.418	1.471
Aliphatic EPH	51244754.000	360.000	3.140	22.053	142346.539	134589.790	-5.763