

Continuing Calibration Report for SequenceID : FE111124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE051137.D

Aliphatic C9-C12	9844255.000	60.000	3.140	6.760	164070.917	139914.732	-17.265
Aliphatic C12-C16	6592706.000	40.000	6.761	10.193	164817.650	140635.568	-17.195
Aliphatic C16-C21	9666922.000	60.000	10.194	13.553	161115.367	137768.027	-16.947
Aliphatic C21-C28	12623455.000	80.000	13.554	17.209	157793.188	133951.574	-17.799
Aliphatic C28-C40	17912037.000	120.000	17.210	22.059	149266.975	128748.418	-15.937
Aliphatic EPH	56639375.000	360.000	3.140	22.059	157331.597	134589.790	-16.897