

Continuing Calibration Report for SequenceID : FE111124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE051146.D

Aliphatic C9-C12	8992876.000	60.000	3.140	6.760	149881.267	139914.732	-7.123
Aliphatic C12-C16	5971459.000	40.000	6.761	10.193	149286.475	140635.568	-6.151
Aliphatic C16-C21	8722430.000	60.000	10.194	13.553	145373.833	137768.027	-5.521
Aliphatic C21-C28	11370681.000	80.000	13.554	17.209	142133.513	133951.574	-6.108
Aliphatic C28-C40	15214138.000	120.000	17.210	22.059	126784.483	128748.418	1.525
Aliphatic EPH	50271584.000	360.000	3.140	22.059	139643.289	134589.790	-3.755