

Continuing Calibration Report for SequenceID : FF110824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF014987.D

Aromatic C10-C12	6049046.000	40.000	4.496	6.365	151226.150	154569.237	2.163
Aromatic C12-C16	9009441.000	60.000	6.366	8.961	150157.350	153560.844	2.216
Aromatic C16-C21	11005066.000	80.000	8.962	13.378	137563.325	142096.237	3.190
Aromatic C21-C36	23927620.000	180.000	13.379	18.826	132931.222	138041.042	3.702
Aromatic EPH	49991173.000	360.000	4.496	18.826	138864.369	143365.296	3.139