

Continuing Calibration Report for SequenceID : FF111124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF014998.D

Aromatic C10-C12	6299210.000	40.000	4.510	6.375	157480.250	154569.237	-1.883
Aromatic C12-C16	9343627.000	60.000	6.376	9.068	155727.117	153560.844	-1.411
Aromatic C16-C21	11272623.000	80.000	9.069	13.383	140907.788	142096.237	0.836
Aromatic C21-C36	24616511.000	180.000	13.384	18.828	136758.394	138041.042	0.929
Aromatic EPH	51531971.000	360.000	4.510	18.828	143144.364	143365.296	0.154