

Continuing Calibration Report for SequenceID : FG110824AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014712.D

Aliphatic C9-C12	8279567.000	60.000	3.223	6.844	137992.783	131341.150	-5.064
Aliphatic C12-C16	5648242.000	40.000	6.845	10.289	141206.050	135821.940	-3.964
Aliphatic C16-C21	8324306.000	60.000	10.290	13.661	138738.433	141250.653	1.779
Aliphatic C21-C28	10463638.000	80.000	13.662	17.327	130795.475	140581.144	6.961
Aliphatic C28-C40	12316329.000	120.000	17.328	22.264	102636.075	105686.976	2.887
Aliphatic EPH	45032082.000	360.000	3.223	22.264	125089.117	126992.540	1.499