

Continuing Calibration Report for SequenceID : FG110824AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014719.D

Aliphatic C9-C12	8042156.000	60.000	3.223	6.844	134035.933	131341.150	-2.052
Aliphatic C12-C16	5549385.000	40.000	6.845	10.289	138734.625	135821.940	-2.144
Aliphatic C16-C21	8165847.000	60.000	10.290	13.661	136097.450	141250.653	3.648
Aliphatic C21-C28	10243944.000	80.000	13.662	17.327	128049.300	140581.144	8.914
Aliphatic C28-C40	12602267.000	120.000	17.328	22.264	105018.892	105686.976	0.632
Aliphatic EPH	44603599.000	360.000	3.223	22.264	123898.886	126992.540	2.436