

Continuing Calibration Report for SequenceID : FG111124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014723.D

Aliphatic C9-C12	8558501.000	60.000	3.223	6.843	142641.683	131341.150	-8.604
Aliphatic C12-C16	5988048.000	40.000	6.844	10.286	149701.200	135821.940	-10.219
Aliphatic C16-C21	9016323.000	60.000	10.287	13.658	150272.050	141250.653	-6.387
Aliphatic C21-C28	11869037.000	80.000	13.659	17.322	148362.963	140581.144	-5.535
Aliphatic C28-C40	13461646.000	120.000	17.323	22.254	112180.383	105686.976	-6.144
Aliphatic EPH	48893555.000	360.000	3.223	22.254	135815.431	126992.540	-6.948