

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4721 **SAS No :** P4721 **SDG No:** P4721
Sample No : PB164790BS **Datafile:** FC067671.D
Client ID : PB164790BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10.0	5.28	53		(40-140)
Aliphatic C12-C16	6.7	4.52	68		(40-140)
Aliphatic C16-C21	10.0	8.05	80		(40-140)
Aliphatic C21-C28	13.3	11.0	82		(40-140)
Aliphatic C28-C40	20.0	18.9	94		(40-140)