

Analytical Sequence

Client: ENTACT

SDG No.: P4721

Project: North Point

Instrument ID: FID_B

GC Column: RTX-502.2

ID: 0.53 (mm)

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

MEAN SUROGATE RT FROM INITIAL CALIBRATION 8.7908					
EPA SAMPLE NO.	LAB SAMPLE ID	DATE AND TIME ANALYZED	DATAFILE	RT	#
20 PPB GRO STD	20 PPB GRO STD	6 Nov 2024 14:11	FB031104.D	8.792	
VBF1106S2	VBF1106S2	6 Nov 2024 15:18	FB031106.D	8.796	
BSF1106S1	BSF1106S1	6 Nov 2024 15:45	FB031107.D	8.796	
EX-2-TPH-3	P4721-03	6 Nov 2024 17:06	FB031110.D	8.828	
BSF1106S2	BSF1106S2	6 Nov 2024 18:54	FB031114.D	8.796	
20 PPB GRO STD	20 PPB GRO STD	6 Nov 2024 19:21	FB031115.D	8.796	
EX-11-TPH-4	P4721-07	6 Nov 2024 20:42	FB031116.D	8.846	
EX-9-TPH-1	P4721-08	6 Nov 2024 21:09	FB031117.D	8.795	
EX-9-TPH-2	P4721-09	6 Nov 2024 21:36	FB031118.D	8.803	
EX-9-TPH-3	P4721-10	6 Nov 2024 22:02	FB031119.D	8.806	
EX-9-TPH-6	P4721-13	6 Nov 2024 23:23	FB031122.D	8.984	*
20 PPB GRO STD	20 PPB GRO STD	7 Nov 2024 1:10	FB031126.D	8.794	
20 PPB GRO STD	20 PPB GRO STD	7 Nov 2024 10:14	FB031127.D	8.789	
VBF1107S1	VBF1107S1	7 Nov 2024 10:54	FB031128.D	8.792	
VBF1107S2	VBF1107S2	7 Nov 2024 11:21	FB031129.D	8.793	
BSF1107S1	BSF1107S1	7 Nov 2024 11:48	FB031130.D	8.794	
20 PPB GRO STD	20 PPB GRO STD	7 Nov 2024 14:56	FB031137.D	8.796	
EX-9-TPH-8	P4721-15	7 Nov 2024 17:10	FB031140.D	8.838	
EX-2-TPH-1	P4721-01	7 Nov 2024 17:36	FB031141.D	8.803	
EX-2-TPH-2	P4721-02	7 Nov 2024 18:03	FB031142.D	8.847	
EX-11-TPH-1	P4721-04	7 Nov 2024 18:30	FB031143.D	8.822	
EX-11-TPH-2	P4721-05	7 Nov 2024 18:57	FB031144.D	8.833	
EX-11-TPH-3	P4721-06	7 Nov 2024 19:24	FB031145.D	8.835	
EX-9-TPH-4	P4721-11	7 Nov 2024 19:50	FB031146.D	8.806	
20 PPB GRO STD	20 PPB GRO STD	7 Nov 2024 21:11	FB031149.D	8.793	
20 PPB GRO STD	20 PPB GRO STD	8 Nov 2024 2:56	FB031160.D	8.793	
EX-9-TPH-5	P4721-12	8 Nov 2024 10:09	FB031163.D	8.850	
20 PPB GRO STD	20 PPB GRO STD	8 Nov 2024 10:36	FB031164.D	8.793	
EX-9-TPH-7	P4721-14	8 Nov 2024 11:42	FB031166.D	8.803	
20 PPB GRO STD	20 PPB GRO STD	8 Nov 2024 12:46	FB031168.D	8.795	

Column used to flag RT values with an * values outside of QC limits

QC Limits
(± 0.10 minutes)

Lower Limit
8.6908

Upper Limits
8.8908