



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	106	100	106	40 - 160	P	11/07/2024	14:42	LB133344
	Antimony	51.9	50.0	104	40 - 160	P	11/07/2024	14:42	LB133344
	Arsenic	19.9	20.0	99	40 - 160	P	11/07/2024	14:42	LB133344
	Barium	99.8	100	100	40 - 160	P	11/07/2024	14:42	LB133344
	Beryllium	5.88	6.0	98	40 - 160	P	11/07/2024	14:42	LB133344
	Cadmium	6.45	6.0	108	40 - 160	P	11/07/2024	14:42	LB133344
	Calcium	2010	2000	101	40 - 160	P	11/07/2024	14:42	LB133344
	Chromium	9.25	10.0	92	40 - 160	P	11/07/2024	14:42	LB133344
	Cobalt	29.3	30.0	98	40 - 160	P	11/07/2024	14:42	LB133344
	Copper	22.5	20.0	112	40 - 160	P	11/07/2024	14:42	LB133344
	Iron	95.5	100	96	40 - 160	P	11/07/2024	14:42	LB133344
	Lead	11.3	12.0	94	40 - 160	P	11/07/2024	14:42	LB133344
	Magnesium	2000	2000	100	40 - 160	P	11/07/2024	14:42	LB133344
	Manganese	20.5	20.0	102	40 - 160	P	11/07/2024	14:42	LB133344
	Nickel	38.8	40.0	97	40 - 160	P	11/07/2024	14:42	LB133344
	Potassium	1950	2000	98	40 - 160	P	11/07/2024	14:42	LB133344
	Selenium	18.8	20.0	94	40 - 160	P	11/07/2024	14:42	LB133344
	Silver	10.7	10.0	107	40 - 160	P	11/07/2024	14:42	LB133344
	Sodium	1610	2000	81	40 - 160	P	11/07/2024	14:42	LB133344
	Thallium	41.6	40.0	104	40 - 160	P	11/07/2024	14:42	LB133344
	Vanadium	40.4	40.0	101	40 - 160	P	11/07/2024	14:42	LB133344
	Zinc	41.4	40.0	104	40 - 160	P	11/07/2024	14:42	LB133344
CRA	Mercury	0.17	0.2	83	40 - 160	CV	11/07/2024	14:50	LB133334