

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	255000	255000	100	216000	294000	11/07/2024	14:47	LB133344
	Antimony	-0.14			-50	50	11/07/2024	14:47	LB133344
	Arsenic	4.03			-20	20	11/07/2024	14:47	LB133344
	Barium	1.97	6.0	33	-94	106	11/07/2024	14:47	LB133344
	Beryllium	1.25			-6	6	11/07/2024	14:47	LB133344
	Cadmium	6.27	1.0	627	-5	7	11/07/2024	14:47	LB133344
	Calcium	235000	245000	96	208000	282000	11/07/2024	14:47	LB133344
	Chromium	54.7	52.0	105	42	62	11/07/2024	14:47	LB133344
	Cobalt	1.98			-30	30	11/07/2024	14:47	LB133344
	Copper	6.59	2.0	330	-18	22	11/07/2024	14:47	LB133344
	Iron	95800	101000	95	85600	116500	11/07/2024	14:47	LB133344
	Lead	8.91			-12	12	11/07/2024	14:47	LB133344
	Magnesium	257000	255000	101	216000	294000	11/07/2024	14:47	LB133344
	Manganese	4.04	7.0	58	-13	27	11/07/2024	14:47	LB133344
	Nickel	2.14	2.0	107	-38	42	11/07/2024	14:47	LB133344
	Potassium	27.1			0	0	11/07/2024	14:47	LB133344
	Selenium	-18.3			-20	20	11/07/2024	14:47	LB133344
	Silver	0.85			-10	10	11/07/2024	14:47	LB133344
	Sodium	-30.2			0	0	11/07/2024	14:47	LB133344
	Thallium	4.42			-40	40	11/07/2024	14:47	LB133344
	Vanadium	7.40			-40	40	11/07/2024	14:47	LB133344
	Zinc	6.19			-40	40	11/07/2024	14:47	LB133344
ICSAB01	Aluminum	261000	247000	106	209000	285000	11/07/2024	14:51	LB133344
	Antimony	621	618	100	525	711	11/07/2024	14:51	LB133344
	Arsenic	116	104	112	88.4	120	11/07/2024	14:51	LB133344
	Barium	504	537	94	437	637	11/07/2024	14:51	LB133344
	Beryllium	498	495	101	420	570	11/07/2024	14:51	LB133344
	Cadmium	999	972	103	826	1120	11/07/2024	14:51	LB133344
	Calcium	238000	235000	101	199000	271000	11/07/2024	14:51	LB133344
	Chromium	556	542	103	460	624	11/07/2024	14:51	LB133344
	Cobalt	503	476	106	404	548	11/07/2024	14:51	LB133344
	Copper	493	511	96	434	588	11/07/2024	14:51	LB133344
	Iron	98300	99300	99	84400	114500	11/07/2024	14:51	LB133344
	Lead	56.6	49.0	116	37	61	11/07/2024	14:51	LB133344
	Magnesium	260000	248000	105	210000	286000	11/07/2024	14:51	LB133344
	Manganese	484	507	96	430	584	11/07/2024	14:51	LB133344
	Nickel	991	954	104	810	1100	11/07/2024	14:51	LB133344
	Potassium	10.4			0	0	11/07/2024	14:51	LB133344
	Selenium	32.9	46.0	72	26	66	11/07/2024	14:51	LB133344
	Silver	201	201	100	170	232	11/07/2024	14:51	LB133344
	Sodium	-89.4			0	0	11/07/2024	14:51	LB133344
	Thallium	106	108	98	68	148	11/07/2024	14:51	LB133344

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	493	491	100	417	565	11/07/2024	14:51	LB133344
	Zinc	835	952	88	809	1095	11/07/2024	14:51	LB133344