

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Walsh Construction Company II, LLC **SDG No.:** P4722
Contract: WALS01 **Lab Code:** CHEM **Case No.:** P4722 **SAS No.:** P4722
Matrix: Solid **Level:** LOW **Client ID:** JC-701-COMP-01A
Sample ID: P4720-01 **Spiked ID:** P4720-01A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Beryllium	mg/Kg	75 - 125	7.44		0.67		10.4	65		P
Chromium	mg/Kg	75 - 125	26.6		12.2		20.7	69		P
Copper	mg/Kg	75 - 125	34.2		22.0		15.5	78		P
Potassium	mg/Kg	75 - 125	1480		1180		520	59		P
Selenium	mg/Kg	75 - 125	74.8		1.04	U	100	75		P
Silver	mg/Kg	75 - 125	3.33		0.46	J	3.90	74		P
Vanadium	mg/Kg	75 - 125	28.1		17.3		15.5	69		P