

SOP ID : M1311-TCLP-15
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-4
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pipette ID : WC
Tumbler ID : T-1 / T-2
TCLP Filter ID : 114771

Start Prep Date : 11/06/2024 Time : 16:45
End Prep Date : 11/07/2024 Time : 09:35
Combination Ratio : 20
ZHE Cleaning Batch : N/A
Initial Room Temperature: 24 °C
Final Room Temperature: 23 °C
TCLP Technician Signature : *[Signature]*
Supervisor By : *[Signature]*

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP108622
HCL-TCLP,1N	N/A	WP108584
HNO3-TCLP,1N	N/A	WP108585
pH Strips	N/A	W1931,W1934,W2350,W2755
pH Strips	N/A	W1937,W1938,W1939,W1940,W1941,W1942
1 Liter Amber	N/A	23091
120ml Plastic bottle	N/A	21029
1:1 HNO3	MP81119	N/A

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-1 / T-2 checked, 30 rpm. p4718-03 is used for MS-MSD. Particle size reduction is not required.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
11/07/24 11:00	<i>[Signature]</i> Prep Group	<i>[Signature]</i> Analysis Group