



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC SDG No.: P4722
Contract: WALS01 Lab Code: CHEM Case No.: P4722 SAS No.: P4722
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.24	0.2	119	40 - 160	CV	11/12/2024	14:39	LB133407
CRI01	Aluminum	98.5	100	98	40 - 160	P	11/14/2024	14:40	LB133457
	Antimony	47.2	50.0	94	40 - 160	P	11/14/2024	14:40	LB133457
	Arsenic	20.6	20.0	103	40 - 160	P	11/14/2024	14:40	LB133457
	Barium	95.9	100	96	40 - 160	P	11/14/2024	14:40	LB133457
	Beryllium	5.63	6.0	94	40 - 160	P	11/14/2024	14:40	LB133457
	Cadmium	6.51	6.0	108	40 - 160	P	11/14/2024	14:40	LB133457
	Calcium	1930	2000	96	40 - 160	P	11/14/2024	14:40	LB133457
	Chromium	9.55	10.0	96	40 - 160	P	11/14/2024	14:40	LB133457
	Cobalt	27.6	30.0	92	40 - 160	P	11/14/2024	14:40	LB133457
	Copper	21.3	20.0	107	40 - 160	P	11/14/2024	14:40	LB133457
	Iron	93.2	100	93	40 - 160	P	11/14/2024	14:40	LB133457
	Lead	11.8	12.0	98	40 - 160	P	11/14/2024	14:40	LB133457
	Magnesium	1920	2000	96	40 - 160	P	11/14/2024	14:40	LB133457
	Manganese	19.6	20.0	98	40 - 160	P	11/14/2024	14:40	LB133457
	Nickel	36.7	40.0	92	40 - 160	P	11/14/2024	14:40	LB133457
	Potassium	1850	2000	92	40 - 160	P	11/14/2024	14:40	LB133457
	Selenium	17.4	20.0	87	40 - 160	P	11/14/2024	14:40	LB133457
	Silver	9.72	10.0	97	40 - 160	P	11/14/2024	14:40	LB133457
	Sodium	1670	2000	83	40 - 160	P	11/14/2024	14:40	LB133457
	Thallium	37.5	40.0	94	40 - 160	P	11/14/2024	14:40	LB133457
	Vanadium	38.4	40.0	96	40 - 160	P	11/14/2024	14:40	LB133457
	Zinc	41.6	40.0	104	40 - 160	P	11/14/2024	14:40	LB133457



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC

SDG No.: P4722

Contract: WALS01

Lab Code: CHEM

Case No.: P4722

SAS No.: P4722

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	95	40 - 160	CV	11/08/2024	10:28	LB133352
CRI01	Arsenic	20.1	20.0	100	40 - 160	P	11/08/2024	14:20	LB133365
	Barium	95.0	100	95	40 - 160	P	11/08/2024	14:20	LB133365
	Cadmium	6.16	6.0	103	40 - 160	P	11/08/2024	14:20	LB133365
	Chromium	9.24	10.0	92	40 - 160	P	11/08/2024	14:20	LB133365
	Lead	11.8	12.0	98	40 - 160	P	11/08/2024	14:20	LB133365
	Selenium	19.8	20.0	99	40 - 160	P	11/08/2024	14:20	LB133365
	Silver	10.3	10.0	103	40 - 160	P	11/08/2024	14:20	LB133365
CRI01	Arsenic	21.8	20.0	109	40 - 160	P	11/19/2024	20:19	LB133525
	Barium	99.9	100	100	40 - 160	P	11/19/2024	20:19	LB133525
	Cadmium	7.16	6.0	119	40 - 160	P	11/19/2024	20:19	LB133525
	Chromium	10.5	10.0	105	40 - 160	P	11/19/2024	20:19	LB133525
	Lead	11.3	12.0	94	40 - 160	P	11/19/2024	20:19	LB133525
	Selenium	17.2	20.0	86	40 - 160	P	11/19/2024	20:19	LB133525
	Silver	10.3	10.0	103	40 - 160	P	11/19/2024	20:19	LB133525